

http://www.natsca.org

# The Biology Curator

Title: FORMALDEMETER 3

Author(s): Not Listed

Source: Not Listed (1997). FORMALDEMETER 3. The Biology Curator, 9, 12 - 13.

URL: <a href="http://www.natsca.org/article/483">http://www.natsca.org/article/483</a>

NatSCA supports open access publication as part of its mission is to promote and support natural science collections. NatSCA uses the Creative Commons Attribution License (CCAL) <a href="http://creativecommons.org/licenses/by/2.5/">http://creativecommons.org/licenses/by/2.5/</a> for all works we publish. Under CCAL authors retain ownership of the copyright for their article, but authors allow anyone to download, reuse, reprint, modify, distribute, and/or copy articles in NatSCA publications, so long as the original authors and source are cited.

# **Conference Report**

## 3. Fuel Cell Sensors (The Formaldemeter)

Dr. W. J. Criddle, Consultant, School of Pure and Applied Biology, University of Wales, Cardiff

#### Abstract


Fuel cell sensors are electrochemical cells in which a substance is catalytically oxidised giving rise to an electric potential which results in a flow of current in an external circuit. Fuels are commonly readily oxidisable organic substances, formaldehyde being one example and catalysts are usually highly surface active platinum based types. The detailed mechanism is complex and is not completely understood, but the overall system may be represented simply (in the case of formaldehyde) as

$$[O_5H_2O]$$
  
2HCHO  $\rightarrow$  2HCOOH + 2H<sup>+</sup> + 2e-  
HCHO  $\rightarrow$  CO<sub>2</sub> + 2H<sup>+</sup> + 2e-

Fuel cells are extensively used in ethanol (drink driving) monitoring equipment, but can, in theory, be used for any oxidisable species, including inorganic species such as sulphur diozide and carbon monoxide. Most of the current research into fuel cells is not directed towards sensors, but towards their use as power sources, especially in the automotive industry.

## Personal details - Dr. W. J. Criddle

Dr. Criddle has until recently been Senior Lecturer in Analytical Chemistry in the Department of Chemistry, University of Wales, Cardiff. During this time, he has published about 70 papers in the general area of analytical chemistry, and particularly in the area of fuel cell sensors. He is presently Consultant in Analytical Chemistry to the School of Pure and Applied Biology, University of Wales, Cardiff, and to several companies associated with analytical chemistry, including PPM Ltd., manufacturers of a range of formaldehyde monitoring equipment. PhD in Chemistry University of Wales, Cardiff 1960.



### FORMALDEMETER<sup>TM</sup> 3

The NEW easy-to-use Formaldemeter<sup>TM</sup> 3 from PPM Ltd. gives a rapid indication of whether the formaldehyde level is below the Maximum Exposure Limit.

- Uses the well proven Lion electrochemical fuel cell sensor
- · Has improved selectivity
- Stores time to peak readings
- · Stores the peak readings

Research has shown that with the Lion fuel cell, the measured 'time to peak' reading is a function of the gases being analysed. Alcohols, which may be present in the environment where aldehyde is measured, have approximately double the 'time to peak" of formaldehyde 'window' will indicate that an interfering substance is present. This qualitative test will therefore serve to indicate whether a more sophisticated analysis is required. With the use of a microprocessor, calibration and zero pots have been eliminated. The maximum peak reading is retained on the display until the unit is switched off.

# **Applications**

Formaldehyde is one of the most commonly used substances in industry, and occurs in many processes and products such as:

- Medical
- · Pharmaceutical
- Particle boards
- Laminated boards
- · Synthetic resins
- Paint manufacture
- Paper manufacture
- Dye stuffs
- Textile treatments
- Horticulture
- Deodorants

## COSHH

The COSHH regulations based on Maximum Exposure Limits make the Formaldemeter 3 an ideal instrument for obtaining a rapid reading, thus eliminating expensive complex and time-consuming laboratory methods.

# Principle of operation

The formaldehyde vapour when drawn across the fuel cell sensor undergoes catalytic oxidation on the platinum surface. This produces an electrical output directly proportional to the formaldehyde level in the atmosphere.

#### How to use

- Hold the instrument in the atmoshere to be analysed and switch ON.
- 2. Depress the sample buttom.
- The formaldehyde level is displayed in approximately 8 seconds.

#### Calibration

The Formaldemeter will normally hold accurate calibration for several months. An easy-to-use formaldehyde standard is included for periodic checking and recalibration.

### Selectivity

A high degree of selectivity has been incorporated into the fuel cell sensor. However, some substances such as ethanol, propanol and butanol can interfere. By comparing the time to peak with that of the standard it is possible to determine if interfering substances are present and whether more sophisticated analysis is required.

#### Phenol filters

Complete removal of contaminants such as phenol and resorcinol can be achieved by attaching the filters provided to the instruments sampling port. The data shown in Table 2 demonstrates this for the Formaldemeter MK II.

## Table 1. Accuracy

me std FORMALDEMETER MK II (F) versus The Modified NIOSH (N) Method (Formaldehyde Concentration (ppm))

	N	F	N	F	N	F
	0.37	0.4	1.84	1.9	5.07	4.9
	0.37	0.4	1.96	1.9	4.94	5.0
	0.31	0.3	1.93	1.7	5.07	5.1
	0.34	0.3	1.93	1.7	5.15	5.1
ean	0.35	0.35	1.91	1.80	5.05	5.03
dev	0.03	0.05	0.05	0.10	0.10	0.09

Reference: Am. Ind. Hyg. Assoc. J. 46(10): 578-584(1985)

Table 2. Phenol filters

Results at 21.3°C from 0.5g formaldehyde 1<sup>-1</sup> and 80g PHENOL 1<sup>-1</sup> STANDARD (Formaldehyde Concentration 2.7 ppm)

Response (ppm)

1 41 /								
Instr. no.	67	74	97	132				
without filter	53.5	42.1	43.1	37.1				
with filter	2.7	2.8	2.6	2.6				

Reference: BTTG Shirley Technology Centre

# TECHNICAL SPECIFICATIONS

### Sensor:

Electrochemical Fuel Cell

### Calibration:

In Field Formaldehyde impregnated silica gel standard

### Display:

3 digit LCD

## Sampling rate:

0-2 mins depending on previous reading

## Sample volume:

10 cm<sup>3</sup> approx.

# **Detection range:**

0.05-50 ppm

#### Precision:

10% at 2.0 ppm level

## Response time:

Approximately 8 seconds from sampling formaldehyde

#### **Battery life:**

300 field tests approx. (PP3 size alkaline battery)

#### Weight:

Instrument: 270 g Complete Kit: 750 g

#### **Dimensions:**

Instrument:  $150 \times 80 \times 34$  mm Carrying Case:  $266 \times 230 \times 50$  mm

#### Kit contents:

1 instrument incl. battery 1 calibration standard 10 phenol filters 1 handbook

## Accessories:

Calibration Standard Phenol filters (100 pack)

